



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN - 3 1997

Ms. Lillie C. Thomas, M.S.
Executive Director of Quality Assurance
Custom Services International, Inc.
3111 West Post Road
Las Vegas, Nevada 89118

Re: K970791
ExtraWear™, Gentlemen's Choice™, Extra Strength
Latex Condom with Nonoxynol-9
Dated: February 28, 1997
Received: March 4, 1997
Regulatory class: II
21 CFR §884.5310/Product code: 85 LTZ

Dear Ms. Thomas:


We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Good Manufacturing Practice for Medical Devices: General (GMP) regulation (21 CFR Part 820) and that, through periodic GMP inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4616. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,


Lillian Yin, Ph.D.
Director, Division of Reproductive,
Abdominal, Ear, Nose and Throat
and Radiological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Page 1 of 1

510(k) Number: K970791

Device Name: Condom (rubber) Contraceptive 85-LTZ

Indications for Use:

Extra strength lubricated rubber male condom with Nonoxynol-9 for use as a contraceptive and prophylactic.

Additional Statement: "Laboratory tests of physical properties show the ExtraWear condom is stronger than the InnerWear condom. However, the breakage rate during sex has not been tested."

Additional Statement: This condom contains a latex condom and a spermicidal lubricant. The spermicide, Nonoxynol-9, reduces the number of active sperm, thereby decreasing the risk of pregnancy if you lose your erection before withdrawal and some semen spill outside the condom. However the extent of the decreased risk has not been established. This condom should not be used as a substitute for the combined use of a vaginal spermicide and a condom

Date Submitted: February 28, 1997

(PLEASE DO NOT WRITE BELOW THIS LINE--CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Robert D. Matting
(Division Sign-Off)
Division of Reproductive, Abdominal, ENT,
and Radiological Devices

510(k) Number K970791

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter Use ☒
(Optional Format 1-2-96)